

Certificate of Verification

XL3t-69288

Reading No 1
Mode Mining
Time 2015-04-29 14:39
Duration 121.20
Units ppm
Sigma Value 2
Sequence Final
Flags
INV NR Gm 1025
OBJEKT TAFELGEMAELEN
MATERIAL HOLZ METALL FASSUNG
ANALYSESTELLE P1 RAND
MESSAUFBAU STATIV ANLIEGEND
FILTER m30lo30hi30li30
User Login Raquet



	ppm	+/-	Error
Ce	-339.000	+/-	123.159
La	1.000	+/-	56.984
Ba	1.000	+/-	61.127
Sb	1.000	+/-	22.483
Sn	25.643	+/-	11.805
Cd	18.154	+/-	8.655
Ag	6.958	+/-	5.554
Bal	912436.813	+/-	692.924
Nb	29.570	+/-	14.213
Zr	1.000	+/-	58.890
Y	1.000	+/-	12.677
Sr	98.673	+/-	8.756
Rb	1.000	+/-	23.695
Th	1.000	+/-	704.142
Bi	1.000	+/-	775.151
Au	1216.424	+/-	66.538
As	5628.242	+/-	146.412
Se	30.126	+/-	21.519
Pb	17420.057	+/-	271.848
Hg	4636.106	+/-	79.454
Zn	473.565	+/-	32.133
Cu	355.983	+/-	28.306
Ni	1.000	+/-	35.900
Co	1.000	+/-	30.319
Fe	868.996	+/-	75.643
Mn	107.012	+/-	73.257
Cr	1.000	+/-	43.215
V	1.000	+/-	32.721
Ti	1.000	+/-	158.490
Ca	33890.008	+/-	546.466
K	3322.219	+/-	195.877
Al	1.000	+/-	909.317
P	3943.513	+/-	149.897
Si	4053.724	+/-	339.968
Cl	1465.512	+/-	45.851
S	9972.705	+/-	176.704
Mg	1.000	+/-	7977.257

Supervised By: _____

Certificate of Verification

XL3t-69288

Reading No 2
Mode Mining
Time 2015-04-29 14:43
Duration 121.50
Units ppm
Sigma Value 2
Sequence Final
Flags
INV NR Gm 1025
OBJEKT TAFELGEMAELENDE
MATERIAL HOLZ METALL FASSUNG
ANALYSESTELLE P2 gelber ROCK
MESSAUFBAU STATIV ANLIEGEND
FILTER m30lo30hi30li30
User Login Raquet



	ppm	+/-	Error
Ce	-339.000	+/-	221.971
La	1.000	+/-	101.705
Ba	1.000	+/-	156.574
Sb	1.000	+/-	34.194
Sn	9129.115	+/-	206.380
Cd	44.656	+/-	16.014
Ag	13.830	+/-	9.178
Bal	857890.000	+/-	1873.851
Nb	160.162	+/-	50.340
Zr	1.000	+/-	152.457
Y	1.000	+/-	41.397
Sr	82.252	+/-	10.869
Rb	1.000	+/-	81.582
Th	1.000	+/-	2000.055
Bi	1.000	+/-	1877.797
Au	1.000	+/-	44.240
As	12367.761	+/-	376.298
Se	51.085	+/-	36.489
Pb	76853.922	+/-	1473.678
Hg	1.000	+/-	26.694
Zn	120.012	+/-	18.073
Cu	2889.290	+/-	95.027
Ni	1.000	+/-	48.416
Co	44.637	+/-	37.936
Fe	139.341	+/-	72.714
Mn	1.000	+/-	159.887
Cr	1.000	+/-	36.925
V	1.000	+/-	26.077
Ti	1.000	+/-	132.775
Ca	4223.879	+/-	249.858
K	880.890	+/-	210.007
Al	1180.732	+/-	592.936
P	1094.768	+/-	121.278
Si	5332.214	+/-	328.262
Cl	1517.682	+/-	49.104
S	25983.738	+/-	566.295
Mg	1.000	+/-	6165.329

Supervised By: _____